

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|---|--|
|  10390 Accredited to ISO 15189:2012 | Spire Healthcare Limited | |
| | Issue No: 002 Issue date: 10 January 2022 | |
| | Spire Hull & East Riding Hospital Pathology Lowfield Road Anlaby Hull | Contact: Sally Eaglesfield Tel: +44 (0) 1482 659 471 E-Mail: sally.eaglesfield@spirehealthcare.com Website: www.spirehealthcare.com |
| Testing performed at the above address only | | |

Site activities performed away from the locations listed above:

| Location details | Activity | Location code |
|-----------------------------|--------------------|---------------|
| Spire Hull Hospital Theatre | Blood issue fridge | N/A |



10390
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Spire Healthcare Limited
Issue No: 002 Issue date: 10 January 2022

Testing performed at main address only

DETAIL OF ACCREDITATION

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used |
|---------------------------|--|--|
| HUMAN BODY FLUIDS | <u>Haematology examinations for the purposes of clinical diagnosis</u> | In-house documented procedures based on equipment manuals as relevant |
| EDTA Whole Blood | Full Blood Count Haemoglobin White Blood Cell Count Red Blood Cell Count Platelet Count Haematocrit Mean Cell Volume Mean Cell Haemoglobin Mean Cell Haemoglobin Concentration Neutrophil count Lymphocyte count Monocyte count Eosinophil count Basophil count | Horiba Yumizen H550 using impedance, cytochemistry and absorbance, SPS-HM-SOP 1700 |
| Citrated Plasma | Coagulation screen PT/INR APTT/APTT ratio | ACL (Werfen) analyser using nephelometry and SPS-HM-SOP0300 |
| Citrated Whole Blood | Estimation of ESR | Manual Westergren Sedatiner method using SPS-HM-SOP0600 |
| EDTA Whole Blood | Blood film assessment for detection of abnormal cell type, morphology, numbers and white cell differential | Manual staining using Leishman's stain and SPS-HM-SOP0500 |
| EDTA Whole Blood | Sickle cell solubility screen | Microgen Bioproducts S Test kit and SPS-HM-SOP0800 |



10390
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Spire Healthcare Limited
Issue No: 002 Issue date: 10 January 2022

Testing performed at main address only

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used |
|--|---|---|
| <p>HUMAN BODY FLUIDS (cont'd)</p> <p>Serum, Heparin plasma</p> | <p><u>Biochemistry examinations for the purposes of clinical diagnosis</u></p> <p>Quantification of the following:</p> <p>Urea Creatinine Sodium Potassium GGT Total protein Cholesterol Triglycerides HDL Phosphates Albumin ALP ALT Amylase AST Total bilirubin Calcium CK Glucose CRP LDH Magnesium Uric acid</p> <p>Calculation of the following parameters: eGFR Adjusted calcium Globulin LDL Cholesterol</p> | <p>In-house documented procedures based on equipment manuals as relevant</p> <p>Roche Cobas C311 using colorimetric, enzymatic, ISE and immunoturbidometry and SPS-BC-SOP0100</p> |



10390
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Spire Healthcare Limited

Issue No: 002 **Issue date:** 10 January 2022

Testing performed at main address only

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used |
|-----------------------------|--|---|
| HUMAN BODY FLUIDS (cont'd) | <u>Blood Transfusion examinations for the purposes of clinical diagnosis</u> | In-house documented procedures based on equipment manuals as relevant |
| EDTA Whole Blood and plasma | Blood group detecting ABO and Rhesus. Antibody screen and identification detecting: Rh-D, C, E, c, e, Cw Kell, K, k, kpa Jka, Jkb JSa Fya, Fyb Lewis, Lea, Leb P1 M, N, S, s Lua Cob Xga | Manual method using Grifols Gel Cards (Coombs Anti IgG) and manual check by tube method, SPS-BT-SOP0200 |
| EDTA Whole Blood and plasma | Serological Crossmatching | Manual method using Grifols Gel Cards (Coombs Anti IgG) and SPS-BT-SOP0200 |

END