


United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS MEDICAL</p> <p>8415</p> <p>Accredited to ISO 15189:2012</p>	<h2>Oxford University Hospitals NHS Foundation Trust</h2>	
	<p>Issue No: 006 Issue date: 06 April 2021</p>	
<p>Cellular Pathology John Radcliffe Hospital Headley Way Headington Oxford OX3 9DU</p>	<p>Contact: Sharon Roberts-Gant Tel: +44 (0)1865-220494 Fax: +44 (0)1865-220190 E-Mail: sharon.roberts-gant@ouh.nhs.uk Website: www.ouh.nhs.uk</p>	
<p>Testing performed by the Organisation at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>John Radcliffe Hospital Headley Way Headington Oxford OX3 9DU</p> <p style="text-align: right;">Sharon Roberts-Gant (contact details above)</p>	<p>Diagnostic Cytopathology Histopathology (inc. routine & special staining, frozen sections) Immunocytochemistry In-situ hybridisation Mortuary services (inc. body receipt, storage & release)</p>	JRH
<p>Churchill Hospital Old Road Headington Oxford OX3 7LE</p> <p style="text-align: right;">Sharon Roberts-Gant (contact details above)</p>	<p>Slide production, staining & mounting of fine needle aspirates for diagnostic cytology.</p>	CHF

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>Churchill Hospital Old Road Headington Oxford OX3 7LE</p> <p style="text-align: right;">Sharon Roberts-Gant (contact details above)</p>	<p>Body receipt, storage & release</p>	CHB
<p>Horton General Hospital Oxford Road Banbury Oxfordshire OX16 9AL</p> <p style="text-align: right;">Sharon Roberts-Gant (contact details above)</p>	<p>Body receipt, storage & release</p>	HH



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES AND FLUIDS	<u>Diagnostic cytopathology</u>	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	
Ascitic fluid Bile Biliary brushings Breast cyst fluid Bronchial brushes Bronchial lavage Broncho-alveolar lavage Colonic brushings Hydrocele fluid Oesophageal brushings Ovarian cyst fluid Pericardial fluid Peritoneal fluid Pleural fluid Sputum Urine	Examination to identify or exclude cytological abnormalities for the purpose of diagnosis	<u>Preparation/centrifugation</u> using procedures: NG02, NG03 in conjunction with manufacturer's instructions using: ThermoShandon Cytospin4, Heraeus Labofuge 400: NG08, NG40 Cytoblocks Kit NG16, NG25, NG32	JRH
Ascitic fluid Bile Biliary brushings Breast cyst fluid Bronchial brushes Bronchial lavage Broncho-alveolar lavage Colonic brushings Hydrocele fluid Oesophageal brushings Ovarian cyst fluid Pericardial fluid Peritoneal fluid Pleural fluid Sputum Urine	Examination to identify or exclude cytological abnormalities for the purpose of diagnosis	<u>Staining</u> using in-house procedures for (as appropriate): Papanicolaou: NG41 May Grunwald Giemsa: NG20 Quick Diff: NG21 in conjunction with manufacturer's instructions using: Shandon Varistain 24-4: NG02, NG04, NG19	JRH



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES AND FLUIDS cont'd	<u>Diagnostic cytopathology cont'd</u>	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	
Fine needle aspirates from: Lymph nodes Neck mass Thoracic: lung, lymph node, mediastinal and pleural Pancreatic Breast	Examination to identify or exclude cytological abnormalities for the purpose of diagnosis	<u>Staining</u> using in-house procedures for (as appropriate): Papanicolaou: NG41 May Grunwald Giemsa: NG20 Quick Diff: NG21 in conjunction with manufacturer's instructions using: Shandon Varistain 24-4: NG19	JRH CHF
Slides prepared in house from samples listed above	Morphological assessment and interpretation/diagnosis	<u>Microscopy (qualitative analysis)</u> Using procedures: CPSOP27 in conjunction with manufacturer's instructions using microscopes listed above	JRH CHF
HUMAN TISSUES	<u>Histopathology</u>	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	
Tissues: biopsy, excision, resection and post-mortem specimens	Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	<u>Macroanalysis & dissection</u> using: H4, H6, H140, MO03, MO05, MO06, MO07, MO08, MO09, MO10, MO11, MO12, MO13, MO15, MO16 <u>Processing</u> using: Sakura VIP: H159, H273, Shandon Pathcentre: H158 and Leica Peloris III : H288, H301 processors <u>Embedding</u> using: Thermoshandon Histostar & Leica embedding centres, Leica Histocore Arcadia embedding centre : H21	JRH



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>HUMAN TISSUES (cont'd)</p> <p>Formalin Fixed Paraffin wax blocks of processed tissue</p> <p>Fixed, fresh and frozen tissue</p> <p>FFPE sections on glass slides</p>	<p><u>Histopathology</u> (cont'd)</p> <p>Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis</p> <p>Basophilic and eosinophilic tissue structures</p> <p>Special stains to detect the following:</p> <p>Acidic mucins</p> <p>Acidic mucins, neutral mucins, PAS positive substances</p> <p>Alcian blue mucins, PAS positive substances</p> <p>Alcian blue mucins</p> <p>Bile</p> <p>Haemopoietic cells</p> <p>Gram positive & negative microorganisms</p>	<p>Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):</p> <p><u>Sectioning & slide production</u> using: Leica RM2235&SM2400, Microm HM200 Ergostar & Shandon Finesse microtomes Leica Histocore Biocut microtomes: H22</p> <p><u>Routine Haemotoxylin and Eosin (H&E) staining & coverslipping</u> using: Sakura Tissue Tek Prisma staining machine: H23, H24 Sakura Tissue Tek DRS 200E stainer: H272, H289 Sakura Tissue Tek GLC 550 coverslipper: H25 Sakura Tissue Tek Glas g2 automated coverslipper: H274, CPSOP32</p> <p><u>Special staining</u> Manual methods using procedures: H264</p> <p>Alcian blue</p> <p>Alcian Blue Periodic Acid Schiffs (PAS)</p> <p>Diastase Alcian blue PAS</p> <p>Alcian blue haematoxylin for skin</p> <p>Fouchet</p> <p>Giemsa</p> <p>Gram's stain</p>	<p>JRH</p>



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd) FFPE sections on glass slides	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis Special stains to detect the following: Gram positive & negative microorganisms Reticulin fibres, collagen Fungi, <i>Pneumocystis jirovecii</i> Amyloid Fungi, <i>Pneumocystis jirovecii</i> Collagen, Muscle, RBCs Basement membranes (basal lamina Collagen, muscle, RBCs Fibrin, Muscle, Collagen, RBCs HBsAg (Hep B), Elastin, Copper Associated Proteins Mucins, Glycogen, PAS positive substances - fungi Glycogen Collagen Haemosiderin, Ferric iron, Asbestos, Ferringinous bodies Basement membranes (basal lamina), Glycogen, PAS positive substances Copper, Copper-associated protein Mast cells	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant): <u>Special staining</u> Manual methods (cont'd) Gram-twort Gordon & Sweets for reticulin Grocott's Hexamine Silver Highmans Congo red Grocott's Hexamine Silver Haematoxylin Van Gieson Jones Hexamine silver Masson's trichrome MSB trichrome Orcein PAS PAS Diastase Fast green / Sirius red Perls Prussian Blue Renal PAS Rhodanine Toluidine blue for mast cells	JRH



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd) FFPE sections on glass slides	<p><u>Histopathology</u> (cont'd)</p> <p>Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis Special stains to detect the following:</p> <p><i>Helicobacter pylori</i></p> <p>Mineralised bone, Canaliculi & Lacunae in the centre of the traberculae, Bone marrow, Osteoid seams</p> <p>Leprosy bacilli, RBCs</p> <p>Acid/alcohol fast bacilli</p> <p>Acidic mucins, neutral mucins, PAS positive substances</p> <p>Acidic mucins, PAS positive substances</p> <p>Glycogen evidenced by loss of staining after Diastase</p> <p>Mucins, Glycogen, PAS positive substances – fungi</p> <p>Melanin, Argentaffin, Chromaffin, Lipofuscin</p> <p>Elastic, Collagen, RBCs</p> <p>Smooth muscle</p> <p>Epithelium</p>	<p>Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):</p> <p><u>Special staining</u> Manual methods (Cont'd)</p> <p>Toluidine blue for Helicobacter Pylori</p> <p>Von kossa</p> <p>Wade-Fite</p> <p>Ziehl Neelsen</p> <p>Automated methods Sakura Tissue Tek DRS 200E stainer: H23</p> <p>Alcian Blue Periodic Acid Schiffs (PAS)</p> <p>Diastase Alcian blue PAS</p> <p>Diastase PAS</p> <p>PAS</p> <p>Masson Fontana</p> <p>Miller's Elastic van Gieson</p> <p>Actin (1A4)</p> <p>AE1/3 (multi CK)</p>	JRH



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd)	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	<u>Immunocytochemistry</u> Leica Bond III Immunostainer and SOP: H235	
	Immunocytochemistry to detect the following:		
	Alpha fetal protein	AFP	
	Anaplastic lymphoma kinase	ALK1	
	Alpha 1 anti-trypsin	Alpha 1AT	
	Racemase (Prostate ca)	AMACR (racemase)	
	Amyloid A	Amyloid A (MC1)	
	Lymphoma panel	Annexin	
	Androgen receptor	Androgen receptor	
	Apoptosis/Lymphoma	Bcl-2	
	Lymphoma panel	Bcl-6	
	Epithelial cell types	BerEP4	
	Epithelial	Beta catinin	
	Lymphoma panel	C Myc	
	Complement factor	C4d	
	Renal cell CA	Ca1X	
	Ovarian marker	CA125	
	Pancreatic marker	CA19.9	
	C cell Thyroid	Calcitonin	
	Smooth muscle cells	Caldesmon	
	Muscle	Calponin	



8415
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Immunocytochemistry to detect the following: Mesothelial T cell Lymphoma Hogkins lymphoma Monocytes B Cell Langerhans cells T cell B cell Follicular dendritic cells Follicular dendritic cells T cells Hogkins lymphoma Vascular Vascular/stromal T cell T cell Leucocyte common antigen T cell Neuroendocrine	<u>Immunocytochemistry</u> (cont'd) Calretinin CD 7 CD10 CD15 Leu M1 CD163 CD19 CD1a CD2 CD20 (L26) CD21 (I2G9) CD23 (MHM6) CD3 CD30 CD31 (Jc70) CD34 (Qbend) CD4 368 CD43 (DFT1) CD45 (LCA) CD5 CD56	



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Immunocytochemistry to detect the following:	<u>Immunocytochemistry</u> (cont'd)	
	Platelets	CD61 (Y2/51)	
	B Cell	CD79a (JCB117)	
	Lymphoma	CD123	
	T Cell	CD8	
	Ewings sarcoma	CD99 (Mic II)	
	Colonic transcription factor	CDX2	
	Carcino embryonic antigen	CEA	
	Neuroendocrine	Chromogranin (Dak-A3)	
	Epithelial	CK 19 (BA17)	
	Epithelial	CK 34βE12	
	Epithelial	CK14	
	Epithelial (gastro)	CK20	
	Epithelial	CK8/18 (Cam 5.2)	
	GIST	C-Kit (CD117)	
	Cytomegalovirus	CMV (CCH2)	
	Mantle cell lymphoma	Cyclin D1	
	Endothelial	D240 (podoplanin)	
	Hairy cell leukaemia	DBA.44	



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Immunocytochemistry to detect the following:	<u>Immunocytochemistry</u> (cont'd)	
	Striated muscle	Desmin (DE33)	
	GIST	DOG-1	
	Neural cell adhesion	E Cadherin	
	Ebstein Barr virus	EBV (LMP) (CS1-4)	
	Epithelial	EMA (E29)	
	Oestrogen receptor	ER	
	Skin dendritic cells	Factor 13a	
	Vascular	Factor8	
	Epithelial / mesothelial	CK5	
	Epithelial	CK7	
	Hormone	Gastrin	
	Bladder carcinoma	Gata 3	
	Hormone	Glucagon	
	Liver panel	Glutamine Synthetase	
	Liver panel	Glypican 3	
	Human Chorionic gonadotrophin	HCG	
	Hep B virus	Hep b core	
	Hep B virus	He b surface	



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Immunocytochemistry to detect the following: Hepatocyte Smooth muscle Human papilloma virus Herpes simplex virus Immunoglobulin Immunoglobulin B cell Immunoglobulin Immunoglobulin Immunoglobulin B Cell Germ cell Pancreatic Light chain Macrophage Light chain Macrophage Mast cell Melanocytic Proliferation	<u>Immunocytochemistry</u> (cont'd) Hepar 1 HHF35 HPV HSV Type II IgA IgD IgG IgG4 IgM Inhibin Insulin Kappa KP1 (CD68) Lambda Lysozyme Mast Cell Tryptase Melan A Mib1 (Ki 67)	



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Immunocytochemistry to detect the following:	<u>Immunocytochemistry</u> (cont'd)	
	MSI (HNPCC)	MLH 1	
	Epithelial	MNF116	
	Epithelial	MOC/031	
	MSI (HNPCC)	MSH 2	
	MSI (HNPCC)	MSH 6	
	Hogkins	MUM1	
	Myeloid cells	Myeloperoxidase (MPO)	
	Rhabdomyosarcoma	MYF4	
	Rhabdomyosarcoma	Myo D1	
	Smooth muscle	Myoglobin	
	Smooth muscle	Myosin (Smooth muscle)	
	Neuroblastoma	NB84a	
	Neurofilament	Neurofilament (2F11)	
	Melanocytic	NK1C3 /CD63	
	Neurone specific enolase	NSE	
	Seminoma	Oct 3/4	
	Lymphoma	Oct 2	
	Oncogene	P53 (DO7)	



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Immunocytochemistry to detect the following:	<u>Immunocytochemistry</u> (cont'd)	
	Mesothelial	P63	
	Pancreatic	Pancreatic PP	
	Parvo virus	Parvo Virus	
	Pneumocystis carini	PCP (3F6)	
	Macrophage	PGM1 (CD68)	
	Seminoma	PLAP	
	Membranous nephropathy	PLA2R	
	MSI (HNPCC)	PMS2	
	Progesterone receptor	PR Pgr 12	
	Prostate specific antigen	PSA Poly	
	Prostatic alkaline phosphatase	PSAP	
	Renal cell carcinoma	Renal Cell CC	
	Red blood cells	Ret40f	
	Neural / melanocytic	S100	
	5 Hydroxytryptamine	Serotonin	
	Pancreatic	Somatostatin	
	Spirochaete	Spirochaete	
	Simian aids virus (Renal)	SV 40	



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Immunocytochemistry to detect the following:	<u>Immunocytochemistry</u> (cont'd)	
	Neuroendocrine	Synaptophysin	
	Lymphoma lymphoblastic	Tdt (339)	
	Mesothelial	Thrombomodulin	
	Thyroid	Thyroglobulin	
	Thyroid transcription factor	TTF1	
	Mesenchyme cells	Vimentin (V9)	
	Plasma cells	Vs38	
	Wilms tumour	Wilms WT1	
		<u>Immunocytochemistry</u> Automated using Ventana Benchmark ULTRA Testing Platform SOP H295 Use of the Ventana BenchMark ULTRA staining machine and antibodies as appropriate:	
	Her 2 oncogene	HER-2	
	Helicobacter pylori	H Pylori	
	HHV-8	HHV-8	
	Dysplastic epithelia	P16	



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Immunocytochemistry to detect the following: Anaplastic lymphoma kinase (ALK) B7-H1	<u>Immunocytochemistry</u> (cont'd) ALK D5F3 LM02 PD-L1 TFE3 Adenovirus Leica Bond III Immunostainer Platform and SOP H236 BAP-1 Calreticulin (CALR) CD14 Fumarate hydratase NKX3.1 Parafibromin PSMA SDHB THSD7A ROS1 SOX10 PAX8 MRQ-50 P40 Napsin A PHOX2B Cathepsin K Renin	



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Breast and Gastrointestinal tissues	Human Epidermal growth factor Receptor 2 (HER2) In-situ hybridisation to detect the following: Epstein Barr virus Lambda light chain Kappa light chain	Gene amplification using Ventana Immuno Platform, Dual colour, dual hapten brightfield in situ hybridisation (DDISH) and procedure <u>SOP H295</u> <u>In-situ hybridisation</u> Automated using procedures: H236 and Leica Bond III and probes as appropriate: EBER Lambda Kappa	JRH



8415
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd)	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
FFPE sections on glass slides	Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Immunohistochemistry Manual method for renal biopsies SOP H255 renal IMP method	
	Complement factor	C1Q	
	Complement factor	C3	
	Immunoglobulin	IgA	
	Immunoglobulin	IgG	
	Immunoglobulin	IgM	
		Immunofluorescence Manual method for skin and renal biopsies SOP H40 Immunofluorescence cutting and staining	
	Complement factor (fluorescent)	C1Q FITC	
	Fibrinogen	Fibrinogen FITC	
	Immunoglobulin (fluorescent)	IgA FITC	
	Immunoglobulin (fluorescent)	IgG FITC	
	Immunoglobulin (fluorescent)	IgM FITC	
	Light chain	Kappa FITC	
	Light chain	Lambda FITC	
	Complement factor (fluorescent)	C3 FITC	



8415
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd)	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	JRH
Stained slides prepared in house from samples listed above	Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis		
Fresh tissue	Morphological assessment and interpretation/diagnosis	<u>Microscopy (qualitative analysis)</u> Using procedures: CPSOP27 and microscopes listed above	
Frozen sections of fresh tissue	Frozen section examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	<u>Cryotomy</u> Using procedures: H10, H201 and Leica CM1850 UV & CM1950 cryostats	
Frozen sections of fresh tissue	Basophilic and eosinophilic tissue structures	<u>Routine H&E staining</u> Manual methods using procedures: H23, H289, H171, H179	
Frozen sections of fresh tissue	Special stains to detect the following:	<u>Special staining</u> Manual methods using procedures: H26, H264 and stains as appropriate:	
	Ganglion cells	Acetylcholinesterase	
	Neutral lipids	Oil red o	
	Immunocytochemistry to detect the following:	<u>Immunocytochemistry</u> Manual method for skin and renal biopsies SOP H40	
	Complement factor (fluorescent)	C1Q FITC	
	Complement factor (fluorescent)	C3 FITC	
	Immunoglobulin fluorescent	IgA FITC	
	Immunoglobulin fluorescent	IgG FITC	
	Immunoglobulin Fluorescent	IgM FITC	



8415

Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Oxford University Hospitals NHS Foundation Trust
Issue No:006 Issue date: 06 April 2021

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES (cont'd)	<u>Histopathology</u> (cont'd) Examination to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Macroscopic and Microscopic examination using in house procedures in conjunction with manufacturer's instructions for the following methods (where relevant):	
Frozen sections of fresh tissue	Immunocytochemistry to detect the following: Light chain Light chain fluorescent Fibrinogen	<u>Immunocytochemistry cont'd</u> Manual method for skin and renal biopsies SOP H40 Kappa FITC Lambda FITC <u>Immunocytochemistry</u> Manual method for skin biopsies SOP H40 Fibrinogen FITC	JRH
Slides prepared in house as above	Morphological assessment and interpretation/diagnosis	<u>Microscopy (qualitative analysis)</u> Using procedures: CPSOP27 and microscopes listed above	JRH
Digital images from slides prepared in house as above	Morphological assessment and interpretation/diagnosis	Using digital pathology (slide scanners and viewing software): Philips Intellisite Digital Pathology Suite, H307 Use and Maintenance of Philips™ Slide Scanners: <ul style="list-style-type: none"> • Philips UFS Ultra Fast Whole Slide Imaging Scanner • Philips UVS Ultra Versatile Whole Slide Imaging scanner • Philips IMS Image Management and Storage System Slide Scanning H307_Use_and_Maintenance_of_Philips_Slide_Scanners Slide Reporting IT05 Histodata entry and authorisation using approved equipment	
END			