

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 9012 Accredited to ISO 15189:2012	Spire Healthcare Limited	
	Issue No: 002 Issue date: 06 December 2019	
	Spire Bristol Hospital The Glen Redland Hill Durdham Road Bristol BS6 6UT	Contact: Mark Galta Tel: +44 (0) 117 974 6462 E-Mail: mark.galta@spirehealthcare.com
Testing performed at the above address only		

Site activities performed away from the locations listed above:

Location details	Activity
Theatre level 1, Spire Bristol Hospital	Blood issue fridge

DETAIL OF ACCREDITATION



9012
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Spire Healthcare Limited

Issue No: 002 Issue date: 06 December 2019

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS	<u>Haematology examinations for the purposes of clinical diagnosis</u>	In-house documented procedures based on equipment manuals as relevant
Blood	Full Blood Count Haemoglobin White Blood Cell Count Red Blood Cell Count Platelet Count Haematocrit Mean Cell Volume Mean Cell Haemoglobin Mean Cell Haemoglobin Concentration Neutrophil count Lymphocyte count Monocyte count Eosinophil count Basophil count	Horiba Pentra 80 using impedance, cytochemistry and absorbance, SPS-HM-SOP0100
Plasma	Coagulation screen PT APTT	Instrument Laboratories ACL 9000 coagulometer and SPS-HM-SOP0300
Blood	Estimation of ESR	Manual Seditainer method (rate of sedimentation) using SPS-HM-SOP0600
Blood	Blood film assessment for detection of abnormal cell type, morphology, numbers and white cell differential	Manual staining using Speedy-Diff kit May Grunwald/Giemsa stain and SPS-HM-SOP0500



9012
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Spire Healthcare Limited

Issue No: 002 Issue date: 06 December 2019

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>HUMAN BODY FLUIDS (cont'd)</p> <p>Serum, plasma</p>	<p><u>Biochemistry examinations for the purposes of clinical diagnosis</u></p> <p>Quantification of the following:</p> <p>Urea Creatinine Sodium Potassium GGT Total protein Cholesterol Triglycerides HDL Phosphates Albumin ALP ALT Amylase AST Total bilirubin Calcium CK Glucose CRP</p> <p>LDH Bicarbonate</p> <p>Calculation of the following parameters: eGFR Adjusted calcium Globulin HDL cholesterol Non-HDL cholesterol Cholesterol:HDL ratio</p>	<p>Roche Cobas C311 using colorimetric, enzymatic, immunoturbidometry and SPS-BC-SOP0100</p> <p>Kinetic Urease Kinetic Jaffe Ion selective electrode Ion selective electrode Enzymatic Colourimetric Enzymatic colourimetric Enzymatic colourimetric Enzymatic colouri Moylbdate UV Colourimetric Colourimetric TRIS buffer rate reaction Enzymatic colourimetric NADH rate rection Diazo method Colourimetric UV test rate reaction Hexokinase rate reaction Particle enhanced immunoturmidometry UV test rate reaction Absorbance</p>



9012
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Spire Healthcare Limited
Issue No: 002 Issue date: 06 December 2019

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS (cont'd)	<p><u>Blood Transfusion examinations for the purposes of clinical diagnosis</u></p> <p>Blood group detecting ABO and Rhesus.</p> <p>Antibody screening and identification: Rh-D, C, c, e, cw Kell, K, k, kpa, kpd JK-a, JK-b Lewis, LEa, LEb P1 M, N, F, f Lutheran a, b XGa</p> <p>Serological Crossmatching</p>	<p>In-house documented procedures based on equipment manuals as relevant</p> <p>Manual method using Diamed Gel Cards(ABO/O+ Rev,grp, Coombs Anti IgG) and SPS-BT-SOP0200</p> <p>Manual method using Diamed Gel Cards (ID-DiaCell ABO/I-II-III 3 cell screening kit) and SPS-BT-SOP0200</p>
END		