

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>9740</b> Accredited to ISO 15189:2012</p>	<h3>Spire Healthcare Limited</h3> <p><b>Issue No:</b> 003    <b>Issue date:</b> 23 September 2020</p>	
	<p><b>Pathology Department</b> Spire Gatwick Park Hospital Povey Cross Road Horley RH4 0BB United Kingdom</p>	<p><b>Contact:</b> Juliette Simon <b>Tel:</b> +44 (0) 1293 778920 <b>E-Mail:</b> juliette.simon@spirehealthcare.com <b>Website:</b> www.spirehealthcare.com</p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS	<u>Haematology examinations for the purposes of clinical diagnosis</u>	In-house documented procedures based on equipment manuals as relevant
Blood	Full Blood Count Haemoglobin White Blood Cell Count Red Blood Cell Count Platelet Count Haematocrit Mean Cell Volume Mean Cell Haemoglobin Mean Cell Haemoglobin Concentration Neutrophil count Lymphocyte count Monocyte count Eosinophil count Basophil count	Horiba Pentra 80 using impedance, cytochemistry and absorbance, SPS-HM-SOP0100
Plasma	Coagulation screen PT APTT	Stago Start analyser using viscosity based detection and SPS-HM-SOP0400
Blood	Estimation of ESR	Manual Westergren Sedatiner method using SPS-HM-SOP0600
Blood	Blood film assessment for detection of abnormal cell type, morphology, numbers and white cell differential	Manual staining using Quick-Diff kit stain and SPS-HM-SOP0500



9740  
Accredited to  
ISO 15189:2012

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Spire Healthcare Limited**  
**Issue No: 003 Issue date: 23 September 2020**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>HUMAN BODY FLUIDS (cont'd)</p> <p>Serum, plasma</p>	<p><u>Biochemistry examinations for the purposes of clinical diagnosis</u></p> <p>Quantification of the following:</p> <p>Urea Creatinine Sodium Potassium GGT Total protein Cholesterol Triglycerides HDL Phosphates Albumin ALP ALT Amylase AST Total bilirubin Calcium CK Glucose CRP LDH</p> <p>Calculation of the following parameters: eGFR Adjusted calcium Globulin LDL Cholesterol</p>	<p>In-house documented procedures based on equipment manuals as relevant</p> <p>Roche Cobas C311 using colorimetric, enzymatic, ISE and immunoturbidometry and SPS-BC-SOP0100</p>



9740  
Accredited to  
ISO 15189:2012

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Spire Healthcare Limited**  
**Issue No: 003 Issue date: 23 September 2020**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS (cont'd)	<u>Blood Transfusion examinations for the purposes of clinical diagnosis</u>	In-house documented procedures based on equipment manuals as relevant
Blood and plasma	Blood group detecting ABO and RhD.  Antibody screen and identification detecting: D, C, E, e, c, Cw, K, k, Kpa, Jsa, Fya, Fyb, Jka, Jkb, Lea, Leb, P1, M, N, S, s, Lua, Cob, Xga	Automated method using Grifols Eflexis analyser Manual method using Grifols gel cards SOP SPS-BT-SOP0200
Blood and plasma	Direct antiglobulin test  Serological Crossmatching	Automated method using Eflexis analyser and manual method using Grifols gel cards SOP-BT-SOP0200  Manual method using Grifols gel cards and SPS-BT-SOP0200  Electronic issue of blood by calculation, SPS-BT-SOP-0200, SPS-BT-WI0225, SPS-IT-SOP0600
END		