


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>4001</b></p> <p>Accredited to ISO 17034:2016</p>	<b>Starna Scientific Ltd</b>	
	<b>Issue No: 011</b>	<b>Issue date: 12 June 2018</b>
	<b>52/54 Fowler Road</b> <b>Hainault</b> <b>Essex</b> <b>IG6 3UT</b>	<b>Contact: John Hammond</b> <b>Tel: +44 (0) 20 8500 1264</b> <b>Fax: +44 (0) 20 8500 1955</b> <b>E-Mail: tech@starna.com</b> <b>Website: www.starna.com</b>
<b>Reference material production at the above address</b>		

### DETAIL OF ACCREDITATION

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
<u>Reference Materials with Optical Properties</u>  Holmium Oxide Solutions Didymium Solutions Samarium Solutions Rare Earth Oxide Solutions Inorganic Salt Stray Light Solutions Holmium Glass Filters Didymium Glass Filters Organic Matrix Solutions	Wavelength (at specific spectral bandwidths)	Measurement by a single, primary, definitive method at Starna Scientific (UV/Vis/NIR Spectrometry)	CRM           CRM
Potassium Dichromate Solutions Neutral Density Filters Nicotinic Acid Organic Matrix Solutions	Visible, Ultraviolet and NIR Absorbance/Transmittance (at specific wavelengths)	Measurement by a single, primary, definitive method at Starna Scientific (UV/Vis/NIR Spectrometry)	CRM    CRM
Neutral Density Metal-on-Quartz filters			CRM
Polystyrene film	Wavelength	Measurement by a single, primary, definitive method at Starna Scientific (IR Spectrometry)	CRM
Other UV/Vis/NIR reference solutions (The organization holds a flexible scope for the certification of suitable materials)	Absorbances are part of the certification under the flexible scope	Measurement by a single, primary, definitive method at Starna Scientific (UV/Vis/NIR Spectrometry)	CRM



4004  
Accredited to  
ISO 17034:2016

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Starna Scientific Ltd**  
**Issue No: 011 Issue date: 12 June 2018**

Reference material certification performed at main address only

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
Other UV/Vis/NIR reference solutions (The organisation holds a flexible scope for the certification of suitable materials)	Relative Fluorescence intensities are part of the Certification under the flexible scope	Measurement by a single, primary, definitive method at Starna Scientific (UV/Vis/NIR Spectrometry)	CRM and RM
END			

**\*Type**

CRM = Certified Reference Material(s)

RM = Reference Material(s)

Refer to ISO 17034 for full definitions