


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS REFERENCE MATERIALS</p> <p>4425</p> <p>Accredited to ISO 17034:2016</p>	<p>Jaytee Biosciences Ltd trading as kromega</p> <p>Issue No: 007 Issue date: 04 February 2020</p>	
	<p>The Boulevard Altira Business Park Herne Bay Kent CT6 6GZ</p>	<p>Contact: Mr Paul Boother Tel: +44 (0)1227 265333 Fax: +44 (0)1227 265331 E-Mail: paul.boother@jaytee.com Website: www.jaytee.com</p>
<p>Reference material production at the above address</p>		

Flexible Scope

The reference material producer is recognised as competent to modify, develop and produce any reference material within the scope of the areas of competence covered by the general scope, and according to and described in the controlled company confidential procedures. The exhaustive list of reference materials covered under accreditation is maintained by, and available from, the reference material producer.

Information about flexible scopes of accreditation is available in UKAS document GEN 4.

DETAIL OF ACCREDITATION

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
<p><u>Reference Materials with Optical Properties</u></p> <p>Holmium Oxide Solutions Imidazole Solutions Rare Earth Solutions Didymium Oxide Solutions</p>	<p>Wavelength (at specific spectral bandwidths)</p>	<p>Measurement by single, primary, definitive methods at Jaytee (UV/Vis Spectrometry)</p>	<p>CRM</p>
<p>Caffeine Solutions Potassium Dichromate Solutions</p>	<p>Visible/Ultraviolet Absorbance (at specific wavelengths)</p>		<p>CRM</p>
<p>Potassium Chloride Solutions</p>	<p>Ultraviolet Absorbance (for stray light)</p>		<p>CRM</p>
<p>Toluene Solutions</p>	<p>Ultraviolet Absorbance (for Resolution)</p>		<p>CRM</p>
<p>Other UV/Vis reference solutions (The organization holds a flexible scope for the certification of suitable materials)</p>	<p>Absorbances are part of the certification under the flexible scope</p>		<p>CRM</p>



4425

Accredited to
ISO 17034:2016

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Jaytee Biosciences Ltd trading as kromega
Issue No: 007 Issue date: 04 February 2020

Reference material production performed by the locations specified

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
<u>Reference Materials for GC Detectors</u>			
Alkane Mixtures in Hexane	Linearity of GC Detector Response	Measurement by single, primary, definitive method at Jaytee (GC-FID and GC-MS)	CRM
Other organic compounds in organic solvent mixtures (The organization holds a flexible scope for the certification of suitable materials)	Linearity of GC detector responses		CRM
END			

***Type**

CRM = Certified Reference Material(s)

RM = Reference Material(s)

Refer to ISO 17034 for full definitions