

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>5710</b> Accredited to ISO 17034:2016</p>	<b>EffecTech Limited</b>	
	<b>Issue No: 001    Issue date: 06 October 2020</b>	
Dove House Dove Fields Uttoxeter Staffordshire ST14 8HU	<b>Contact: Dr Gavin Squire</b> <b>Tel: +44 (0)1889 569229</b> <b>E-Mail: gavin.squire@effectech.co.uk</b> <b>Website: www.effectech.co.uk</b>	
<b>Reference material production at the above address</b>		

### DETAIL OF ACCREDITATION

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)	
SYNTHETIC NATURAL GAS MIXTURES	amount fraction	(% mol/mol)	<b>In-house method TM001/UT</b>  Calibration of certified reference materials (CRM) by ISO 6143:2001 using gas chromatography.	CRM
	nitrogen	(0.1 to 22)		
	carbon dioxide	(0.05 to 15)		
	methane	(34 to 100)		
	ethane	(0.1 to 35)		
	propane	(0.05 to 15)		
	iso-butane	(0.01 to 2)		
	n-butane	(0.01 to 2)		
	neo-pentane	(0.002 to 0.35)		
	iso-pentane	(0.005 to 0.35)		
	n-pentane	(0.005 to 0.35)		
	n-hexane	(0.001 to 0.35)		
	2-methylpentane	(0.001 to 0.35)		
	3-methylpentane	(0.001 to 0.35)		
	2,2-dimethylbutane	(0.001 to 0.35)		
	benzene	(0.001 to 0.2)		
	cyclohexane	(0.001 to 0.2)		
	n-heptane	(0.001 to 0.2)		
	toluene	(0.001 to 0.1)		
	methylcyclohexane	(0.001 to 0.1)		
n-octane	(0.0005 to 0.05)			
n-nonane	(0.0005 to 0.02)			
n-decane	(0.0005 to 0.005)			
helium	(0.005 to 0.2)			
hydrogen	(0.005 to 0.2)			



5710

Accredited to  
ISO 17034:2016

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**EffecTech Limited**

**Issue No:** 001 **Issue date:** 06 October 2020

Reference material production performed at the main address only

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
SYNTHETIC NATURAL GAS MIXTURES (cont'd)	<p>amount fraction (% mol/mol)</p> <p>oxygen (0.001 to 22.5)</p> <p>superior calorific value molar basis (kJ.mol<sup>-1</sup>) mass basis (MJ.kg<sup>-1</sup>) volume basis (MJ.m<sup>-3</sup>)</p> <p>inferior calorific value molar basis (kJ.mol<sup>-1</sup>) mass basis (MJ.kg<sup>-1</sup>) volume basis (MJ.m<sup>-3</sup>)</p> <p>relative density</p> <p>density (kg.m<sup>-3</sup>)</p> <p>superior Wobbe index (MJ.m<sup>-3</sup>)</p> <p>inferior Wobbe index (MJ.m<sup>-3</sup>)</p> <p>molar mass (kg.kmol<sup>-1</sup>)</p> <p>compression factor</p> <p>gross calorific value molar basis (kJ.mol<sup>-1</sup>) mass basis (MJ.kg<sup>-1</sup>) volume basis (MJ.m<sup>-3</sup>)</p> <p>net calorific value molar basis (kJ.mol<sup>-1</sup>) mass basis (MJ.kg<sup>-1</sup>) volume basis (MJ.m<sup>-3</sup>)</p> <p>relative density</p> <p>density (kg.m<sup>-3</sup>)</p> <p>gross Wobbe index (MJ.m<sup>-3</sup>)</p> <p>net Wobbe index (MJ.m<sup>-3</sup>)</p> <p>molar mass (kg.kmol<sup>-1</sup>)</p> <p>compression factor</p>	<p><b>In-house method TM026/UT</b></p> <p>Calibration of certified reference materials (CRM) by ISO 12963:2017 using galvanic fuel cell sensors</p> <p>Values calculated by <b>ISO 6976:1995</b> (including amendment No 1, May 1998) on a <i>real</i> or <i>ideal</i> gas basis assuming mixture is dry (free from water)</p> <p>Combustion properties can be expressed in units of the Joule (J) or in kilowatt hours (kWh)</p> <p>Values calculated by <b>ISO 6976:2016</b> on a <i>real</i> or <i>ideal</i> gas basis assuming mixture is dry (free from water)</p> <p>Combustion properties can be expressed in units of the Joule (J) or in kilowatt hours (kWh)</p>	<p>CRM</p>



5710

Accredited to  
ISO 17034:2016

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**EffecTech Limited**

**Issue No:** 001 **Issue date:** 06 October 2020

Reference material production performed at the main address only

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
SYNTHETIC NATURAL GAS MIXTURES (cont'd)	gross heating value net heating value relative density compressibility factor	Values calculated by methods given in <b>GPA 2172-09</b> (2009) using data tables from <b>GPA 2145-09</b>	CRM
	gross heating value net heating value relative density density compressibility factor	Values calculated by methods given in <b>ASTM D3588-98</b> (2011) using data tables from <b>GPA 2145-09</b>	
SULPHUR GAS MIXTURES	amount fraction (µmol/mol)	<b>In-house method TM002/UT</b>  Calibration of certified reference materials (CRM) using gas chromatography with sulphur chemiluminescence detection (GC-SCD)	CRM
	hydrogen sulphide (0.2 to 10)		
	carbonyl sulphide (0.2 to 10)		
	methanethiol (methyl mercaptan) (0.2 to 10)		
	ethanethiol (ethyl mercaptan) (0.2 to 10)		
	dimethyl sulphide (0.2 to 10)		
	1-propanethiol (n-propyl mercaptan) (0.2 to 10)		
	2-propanethiol (iso-propyl mercaptan) (0.2 to 10)		
	ethyl methyl sulphide (methyl ethyl sulphide) (0.2 to 10)		
	1-butanethiol (n-butyl mercaptan) (0.2 to 10)		
	2-methyl-2-propanethiol (tert-butyl mercaptan) (0.2 to 10)		
	2-methyl-1-propanethiol (iso-butyl mercaptan) (0.2 to 10)		
	1-methyl-1-propanethiol (sec-butyl mercaptan) (0.2 to 10)		
	diethyl sulphide (0.2 to 10)		
	n-hexyl mercaptan (0.2 to 10)		
tetrahydrothiophene (THT) (0.2 to 10)			



5710  
Accredited to  
ISO 17034:2016

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**EffecTech Limited**  
**Issue No:** 001    **Issue date:** 06 October 2020

Reference material production performed at the main address only

Matrix / Artefact	Property Value(s) / Identity / Characterisation Range	Characterisation Procedure / Technique	Type* (CRM / RM)
BINARY EMISSION GAS MIXTURES	amount fraction (% mol/mol)	<b>In-house method TM014</b>  Calibration of certified reference materials (CRM) by ISO 12963:2017 using dynamically generated reference gases in accordance with ISO 6145 Part 7: Thermal Mass Flow Controllers	CRM
	carbon dioxide (0.1 to 15) in nitrogen or synthetic air		
	oxygen (0.5 to 25) in nitrogen		
	methane (0.1 to 5) in nitrogen		
	methane (0.1 to 2.5) in synthetic air		
	amount fraction (µmol/mol)		
	carbon monoxide (10 to 1000) in nitrogen or synthetic air		
	nitric oxide (10 to 600) in nitrogen		
	nitrogen dioxide (5 to 500) in synthetic air		
	sulphur dioxide (10 to 1000) in nitrogen or synthetic air		
	amount fraction (% mol/mol)	<b>In-house method TM026/UT</b>  Calibration of certified reference materials (CRM) by ISO 12963:2017 using galvanic fuel cell sensors	
	oxygen (0.001 to 22.5) in nitrogen		
END			

**\* Type**

CRM = Certified Reference Material(s)  
RM = Reference Material(s)

Refer to ISO 17034 for full definitions