


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 22138 Accredited to ISO/IEC 17025:2017	Moorcroft Controls Limited	
	Issue No: 002	Issue date: 25 August 2021
	Unit 2, The Green Hawks Green Lane Cannock WS11 7LG	Contact: Mr Andrew Bilbrough Tel: +44 (0)1543 509020 E-Mail: andrew.bilbrough@mcontrols.io Website: www.mcontrols.io
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Unit 2, The Green Hawks Green Lane Cannock WS11 7LG	Local contact Mr Andrew Bilbrough Tel: +44 (0)1543 509020 Email: andrew.bilbrough@mcontrols.io	Electrical / Pressure	Lab

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Electrical / Pressure	Site



22138
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Moorcroft Controls Limited
Issue No: 002 Issue date: 25 August 2021

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
Calibrations are performed as a direct comparison against a reference standard unless otherwise stated.				
ELECTRICAL				
DC Resistance	0 Ω to 40 Ω 100 Ω to 400 Ω 400 Ω to 4000 Ω	0.040 Ω 0.10 Ω 1.2 Ω		Site and Lab
DC Voltage Measurement	0 mV to 50 mV 50 mV to 1 V 1V to 10 V	0.05 mV 0.15 mV 1.5 mV		Site and Lab
Generation	0 mV to 100 mV 100 mV to 1 V 1V to 10 V	0.20 mV 0.35 mV 1.5 mV		Site and Lab
DC Current Measurement	4 mA to 20 mA	0.20 mA		Site and Lab
Generation	4 mA to 20 mA	0.20 mA		
Temperature by Electrical Simulation Temperature calibrators, simulators, indicators, recorders, controllers and transmitters, calibration by electrical simulation				
Base metal thermocouple (J, K, T & N) Simulation and Measurement	-50 $^{\circ}\text{C}$ to +1300 $^{\circ}\text{C}$	1.1 $^{\circ}\text{C}$	including cold junction compensation	Site and Lab
Noble metal thermocouple (R & S type) Simulation and Measurement	0 $^{\circ}\text{C}$ to 1700 $^{\circ}\text{C}$	4.1 $^{\circ}\text{C}$	including cold junction compensation	Site and Lab
Pt 100	-190 $^{\circ}\text{C}$ to 0 $^{\circ}\text{C}$ 0 $^{\circ}\text{C}$ to 750 $^{\circ}\text{C}$	0.15 $^{\circ}\text{C}$ 0.30 $^{\circ}\text{C}$	3 and 4 wire measurement	Site and Lab
PRESSURE				
Gas Pressure (Gauge)				
Calibration of pressure indicating instruments and gauges.	-95 kPa to 2000 kPa	1.8 kPa		Site and Lab
END				



22138
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Moorcroft Controls Limited
Issue No: 002 **Issue date:** 25 August 2021

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$