


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS MEDICAL 22083</p> <p>Accredited to ISO 15189:2012</p>	<p>MDNA Life Sciences UK Limited</p> <p>Issue No: 002 Issue date: 06 September 2021</p>	
	<p>MDNA Life Sciences UK Ltd The Biosphere Draymans Way Newcastle Helix Newcastle Upon Tyne NE4 5BX</p>	<p>Contact: Harry Smart E-Mail: h.smart@mdnalifesciences.com Website: www.mdnalifesciences.com</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>HUMAN BODY FLUIDS</p> <p>Combined Nose/throat swab in Viral Transport Medium deactivating or Dry swab</p> <p>Samples self-collected or collected by a contracted customer</p>	<p><u>Molecular Virology examination activities for the purpose of clinical diagnosis</u></p> <p>Detection of SARS-CoV-2 virus RNA (COVID-19); specific gene sequences ORF1ab</p> <p>Detection of SARS-CoV-2 virus RNA (COVID-19); specific gene sequences ORF1ab & S and M gene</p>	<p>In-house documented procedures incorporating manufacturers' instructions & DHSC requirements for private providers of Covid-19 testing, to include:</p> <ul style="list-style-type: none"> • General Population Testing (including fit-to-fly) • Test to Release (day 5) Testing • Day 2/8 Testing <p>Automated nucleic acid extraction and RT PCR using: Qiacube HT extraction and Biorad CFX platforms using:</p> <p>Qiagen Viral RNA extraction kit and PrimerDesign genesig® Real-Time PCR Coronavirus COVID-19 (CE IVD) standard kit</p> <p>SOP WI 1211</p> <p>Qiagen Viral RNA extraction kit and PrimerDesign genesig® Real-Time PCR Coronavirus SARS CoV- 3 Gene kit SOP WI 1211</p>
<p>END</p>		