


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>22200 Accredited to ISO 15189:2012</p>	<h3>DNA Workplace Limited</h3> <p>Issue No: 002 Issue date: 10 September 2021</p>	
	<p>Unit G1 Frome Business Park Manor Road Marston Trading Estate Frome Somerset BA11 4FN</p>	<p>Contact: David Nicholson Tel: +44 0203 9438 371 E-Mail: david@dnaworkplace.com Website: www.dnaworkplace.com</p>
<p>Sampling and testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>HUMAN BODY FLUID</p> <p>Combined Nasopharyngeal & Oropharyngeal (Nose & Throat) Swab</p>	<p><u>Sampling with subsequent testing at a laboratory accredited to ISO 15189:2012 or ISO/IEC 17025:2017 and who is listed on the DHSC Approved List for the testing performed.</u></p> <p>For:</p> <p>General Population Detection of SARS-CoV-2 (COVID-19) virus RNA using molecular methods (inc. PoCT) and Antigen PoCT</p> <p>Test to Release 5 Day Detection of SARS-CoV-2 (COVID-19) virus RNA using molecular methods</p> <p>2 and 8 Day testing for international arrivals Detection of SARS-CoV-2 (COVID-19) virus RNA using molecular methods</p>	<p>In house method based on PHE Guidance for COVID-19 Sampling for testing purposes</p>



22200
Accredited to
ISO 15189:2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

DNA Workplace Limited
Issue No: 002 Issue date: 10 September 2021

Sampling and testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUID	<u>Sampling with subsequent provision of Point of Care Testing (PoCT) in accordance with ISO 22870:2016:</u>	In house method with manufacturer's instructions to meet DHSC minimum requirements for private providers of COVID-19 testing for the general population
Combined Oropharyngeal & Nasopharyngeal (Nose & Throat) Swab	Detection of SARS-CoV-2 (COVID-19) virus RNA specific gene sequence: E	Real-time RT-PCR using Bosch Vivalytic Analyser and SARS-CoV-2 rapid PCR test
Combined Oropharyngeal & Nasopharyngeal (Nose & Throat) Swab	Detection of SARS-CoV-2 (COVID-19) virus Antigen	KIAMGENICS Novel Coronavirus Antigen Detection Kit, chromatographic immunoassay
END		