


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>22223</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Salient Labs Limited trading as Salient Bio</p> <p>Issue No: 001 Issue date: 10 September 2021</p>	
	<p>Salient Labs Limited Unit 506 Cocoa Studios The Biscuit Factory 100 Drummond Road London SE16 4DG</p>	<p>Contact: Marta Ciechońska Tel: +44 (0)203 826 8189 E-Mail: marta@salient.bio Website: www.salient.bio</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>HUMAN TISSUES AND FLUIDS</p> <p>Nasopharyngeal swabs (combined nose and throat)</p> <p>Swabs in virus transport medium (VTM) or inactivating transport medium (ITM)</p>	<p>Molecular examination activities</p> <p>Detection of SARS-CoV-2 virus RNA (COVID-19) from N (nucleocapsid) and E (Envelope) coding regions</p>	<p>Documented in house methods including DHSC requirements for private providers of Covid-19 testing, to include:</p> <ul style="list-style-type: none"> • General Population Testing • Test to Release Testing • Day 2 & 8 International Arrival Testing <p>RNA extraction using Promega Maxwell HT Viral TNA Kit extraction kits, Opentrons 2 and CyBio Felix sampling handling robots</p> <p>Real Time PCR using Nonacus VirPath SARS CoV2 Multiplex qRT PCR kit with N1 and E targets and Analytik Jena qTower or Bio-Rad CFX-96 amplification platforms and Zymo ThermoFisher thermocyclers and in-house method: SOP101-105</p>
<p>END</p>		