


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 22223 Accredited to ISO/IEC 17025:2017	Salient Labs Limited trading as Salient Bio	
	Issue No: 002 Issue date: 10 August 2022	
	Unit 506 Cocoa Studios The Biscuit Factory 100 Drummond Road London SE16 4DG	Contact: Marta Ciechońska Tel: +44 (0)203 826 8189 E-Mail: marta@salient.bio Website: www.salient.bio
Testing performed at the above address only		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN TISSUES AND FLUIDS Nasopharyngeal swabs (combined nose and throat) Swabs in virus transport medium (VTM) or inactivating transport medium (ITM)	Molecular examination activities Detection of SARS-CoV-2 virus RNA (COVID-19) from N (nucleocapsid) and E (Envelope) coding regions	Documented in house methods including DHSC requirements for private providers of Covid-19 testing, to include: <ul style="list-style-type: none">• General Population Testing RNA extraction using Promega Maxwell HT Viral TNA Kit extraction kits, Opentrons 2 and CyBio Felix sampling handling robots Real Time PCR using Nonacus VirPath SARS CoV2 Multiplex qRT PCR kit with N1 and E targets and Analytik Jena qTower or Bio-Rad CFX-96 amplification platforms and in-house method: SOP101-105
END		