


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>21817</b></p> <p>Accredited to ISO/IEC 17025:2017</p>	<p align="center"><b>Thermosense Limited</b></p> <p align="center"><b>Issue No: 003   Issue date: 26 May 2023</b></p>	
	<p><b>Eton House</b> <b>Eton Way North</b> <b>Radcliffe</b> <b>Lancashire</b> <b>M26 2ZT</b></p>	<p><b>Contact: Mr Paul Franklin</b> <b>Tel: +44 (0)162 8531 166</b> <b>E-Mail: paul.franklin@thermosense.co.uk</b> <b>Website: www.thermosense.co.uk</b></p>
<p align="center"><b>Calibration performed at the above address only</b></p>		

### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>TEMPERATURE</b>			
Base Metal thermocouples (with indicators)	-80 °C to 0 °C 0 °C to 100 °C 100 °C to 650 °C 650 °C to 1200 °C	0.40 °C 0.40 °C 0.60 °C 2.8 °C	Calibration in metal block baths
Noble Metal thermocouples Type R and S	0 °C to 100 °C 100 °C to 650 °C 650 °C to 1200 °C	0.47 °C 0.64 °C 2.8 °C	Calibration in metal block baths
Probes and indicators	-80 °C to 0 °C 0 °C to 100 °C 100 °C to 650 °C 650 °C to 1200 °C	0.10 °C 0.20 °C 0.50 °C 2.8 °C	Calibration in metal block baths
4-wire Platinum Resistance Thermometers	-80 °C to 0 °C 0 °C to 100 °C 100 °C to 650 °C	0.10 °C 0.10 °C 0.60 °C	Calibration in metal block baths. 3 wire probes will be subject to additional uncertainty contributions.
Probes/thermocouples with 4-20 mA Transmitters	-80 °C to 0 °C 0 °C to 100 °C 100 °C to 650 °C 650 °C to 1200 °C	0.12 °C 0.20 °C 0.60 °C 2.8 °C	Calibration in metal block baths
Temperature by Electrical Simulation			
PT100 simulators	-200 °C to +850 °C	0.034 °C	
Resistance thermometers	-200 °C to +850 °C	0.034 °C	



21817  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Thermosense Limited**  
**Issue No:** 003 **Issue date:** 26 May 2023

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
Thermocouple Simulation			Including cold junction compensation
Type B	300 to 600 °C 600 to 1820 °C	0.52 °C 0.27 °C	
Type E	-210 to -100 -100 to 1000	0.46 °C 0.24 °C	
Type J	-210 °C to -100 °C -100 °C to 1200 °C	0.46 °C 0.22 °C	
Type K	-210 °C to -100 °C -100 °C to 1300 °C	0.52 °C 0.24 °C	
Type N	-210 °C to -100 °C -100 °C to 1300 °C	0.63 °C 0.23 °C	
Type R	-50 °C to 0 °C 0 °C to 1768 °C	0.48 °C 0.25 °C	
Type S	-50 °C to 0 °C 0 °C to 1768 °C	<b>0.45 °C</b> <b>0.34 °C</b>	
Type T	-210 °C to -100 °C -100 °C to 400 °C	0.48 °C 0.25 °C	
END			



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Thermosense Limited**  
**Issue No:** 003    **Issue date:** 26 May 2023

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$