


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |   |  |
|--|---|--|
| <br><b>UKAS</b><br>CALIBRATION<br><br><b>20481</b><br><br>Accredited to<br>ISO/IEC 17025:2017 | <b>Wessex Power Technology Limited</b>                      |  |
|  | Issue No: 002    Issue date: 15 July 2022                   |  |
|  | Wahn Fried,<br>Winterborne Zelston,<br>Dorset.<br>DT11 9EX. | Contact: Mr Jeremy Rouse<br>Tel: +44 (0)1929 459459<br>Fax: +44 (0)845 520 0304<br>E-Mail: <a href="mailto:Jeremy.rouse@wessexpower.co.uk">Jeremy.rouse@wessexpower.co.uk</a><br>Website: <a href="http://www.wessexpower.co.uk">www.wessexpower.co.uk</a> |

**Calibration performed by the Organisation at the locations specified**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details  | Activity               | Location code |
|---|------------------------|---------------|
| Address<br>Wahn Fried,<br>Winterborne Zelston,<br>Dorset.<br>DT11 9EX.<br><br>Local contact<br>Mr Jeremy Rouse<br>Tel: +44 (0)1929 459459<br>E-Mail: <a href="mailto:Jeremy.rouse@wessexpower.co.uk">Jeremy.rouse@wessexpower.co.uk</a> | Temperature & Humidity | P             |

#### Site activities performed away from the locations listed above:

| Location details   | Activity                        | Location code |
|--|---------------------------------|---------------|
| The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. | Temperature mapping of chambers | S             |



20481

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Wessex Power Technology Limited**

Issue No: 002 Issue date: 15 July 2022

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

| Measured Quantity<br>Instrument or Gauge  | Range  | Expanded Measurement<br>Uncertainty ( $k = 2$ ) | Remarks  | Location<br>Code |
|---|--|---|--|------------------|
| <b>TEMPERATURE</b>  |  |   |  |                  |
| Data recorders/indicators/loggers with sensors  | -80 °C to +250 °C                            | 0.02 °C   | Calibration by comparison in stirred liquid baths  | P                |
|   | -196 °C (liquid nitrogen)                    | 0.06 °C   |  |                  |
| Temperature controlled, incubators, ovens, environmental chambers, fridges/refrigerators, and freezers (inclusive of associated indicators, controllers, and recorders) | -30 °C to +50 °C                             | 0.20 °C   | Single and multipoint time dependent temperature profiling, also referred to as spatial temperature surveying or mapping | S                |
| Air temperature<br>Calibration of temperature loggers and probes in an air chamber  | 0 °C to +60 °C                               | 0.20 °C   | By comparison with reference PRT's in an air chamber   | P                |
| <b>HUMIDITY</b>   |  |   |  |                  |
| Relative Humidity Instruments   | Over the temperature range<br>0 °C to +60 °C |   | By comparison with reference PRT's and a reference Hygrometer  | P & S            |
|   | 0 °C to 60 °C<br>2 %rh to 50 %rh             | 1.00 %rh  |  |                  |
|   | 0 °C to 10 °C<br>50 %rh to 98 %rh            | 2.00 %rh  |  |                  |
|   | 10 °C to 50 °C<br>50 %rh to 98 %rh           | 2.00 %rh  |  |                  |
|   | 50 °C to 60 °C<br>50 %rh to 90 %rh           | 2.00 %rh  |  |                  |
| END   |  |   |  |                  |



20481

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Wessex Power Technology Limited**

**Issue No: 002 Issue date: 15 July 2022**

Calibration performed by the Organisation at the locations specified

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$