

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS TESTING</p> <p>22215</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Inivos Limited, trading as Test Labs</p> <p>Issue No: 001 Issue date: 11 October 2021</p>	
	<p>Unit 22 Coningsby Road Bretton Peterborough PE3 8SB</p>	<p>Contact: Marcus Gordon Tel: +44 (0)845 270 6690 E-Mail: Marcus.gordon@inivos.com Website: www.testlabsuk.com</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>CHEMICAL DISINFECTANTS Including surface disinfectants, handwash, hand rubs</p> <p>Impregnated wipes</p>	<p><u>Microbiological Quantitative Disinfectant Suspension Testing:</u></p> <p>Bacterial quantitative suspension test for food, industrial, domestic and institutional applications</p> <p>Fungicidal or yeasticidal quantitative suspension test for food, industrial, domestic and institutional applications</p> <p>Bactericidal quantitative suspension test for medical applications</p> <p>Fungicidal or yeasticidal quantitative suspension test for medical applications</p> <p><u>Microbiological Quantitative Non-Porous Surface Disinfectant Testing:</u></p> <p>Bactericidal and fungicidal or yeasticidal quantitative non-porous surface test for medical applications (mechanical action employing wipes)</p>	<p>Documented in-house methods based on BS EN standard methods, including customer specified conditions</p> <p>MIC-TP-001 based on BS EN 1276:2019</p> <p>MIC-TP-003 based on BS EN 1650:2019</p> <p>MIC-TP-002 based on BS EN 13727:2012 +A2:2015</p> <p>MIC-TP-005 based on BS EN 13624:2013</p> <p>MIC-TP-007 based on BS EN 16615:2015</p>



22215

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Inivos Limited, trading as Test Labs

Issue No: 001 Issue date: 11 October 2021

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
AUTOMATED ROOM DISINFECTION MACHINES	Bactericidal, sporicidal, fungicidal and yeasticidal evaluation of automated airborne disinfection process	MIC-TP-004 based on BS EN 17272:2020 employing test chamber sizes of 67.5m ³ or 11.3m ³
END		