


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>UKAS</b> MEDICAL <b>22737</b>  Accredited to ISO 15189:2012	<b>Confirm Testing Ltd</b>	
	Issue No: 001    Issue date: 12 October 2021	
	1 Harley Street London W1G 9QD	Contact: Toby Nicol Tel: +44 (0)203 870 3623 E-Mail: ceoffice@confirmtesting.com Website: www.confirmtesting.com
<b>Sampling and testing performed by the Organisation at the location specified below</b>		

### Site activities performed away from the location listed above:

Location details	Activity
Marriot Hotel Bath Road, Hayes UB3 5AN	Sampling for subsequent COVID-19 molecular testing  Sampling and subsequent provision of COVID-19 Point of Care Testing (POCT) in accordance with ISO 22870:2016



22737  
Accredited to  
ISO 15189:2012

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Confirm Testing Ltd**

**Issue No:** 001    **Issue date:** 12 October 2021

Sampling and testing performed by the Organisation at the location specified below

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>HUMAN BODY FLUID</p> <p>Combined Nasopharyngeal &amp; Oropharyngeal (Nose &amp; Throat) Swab</p>	<p><u>Sampling with subsequent testing at a laboratory accredited to ISO 15189:2012 and who is listed on the DHSC Approved List for the testing performed.</u></p> <p>For:</p> <p>General Population Detection of SARS-CoV-2 (COVID-19) virus RNA using molecular methods or antigen POCT</p> <p>Test to Release 5 Day Detection of SARS-CoV-2 (COVID-19) virus RNA using molecular methods</p> <p>2 and 8 Day testing for international arrivals Detection of SARS-CoV-2 (COVID-19) virus RNA using molecular methods</p> <p><u>Sampling and subsequent provision of Point of Care Testing (POCT) in accordance with ISO 22870:2016:</u></p>	<p>In house method based on UKHSA Guidance for COVID-19 Sampling for testing purposes</p> <p>In house method with manufacturer's instructions to meet DHSC minimum requirements for private providers of COVID-19 testing for the general population</p>
<p>Combined Oropharyngeal &amp; Nasopharyngeal (Nose &amp; Throat) Swab</p>	<p>Detection of SARS-CoV-2 (COVID-19) virus antigen</p>	<p>Roche SD Biosensor SARS-CoV-2 Rapid Antigen Test, chromatographic immunoassay</p>

END