


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 23015 Accredited to ISO/IEC 17025:2017	Kordus Consulting Limited	
	Issue No: 001 Issue date: 02 November 2021	
	Ransom Hall Ransom Wood Business Park Southwell Road Mansfield Nottinghamshire NG21 0HJ	Contact: Mr Will Deverill Tel: +44 (0) 333 090 8398 E-Mail: will.deverill@kordus.co.uk Website: www.kordus.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Ransom Hall Ransom Wood Business Park Southwell Road Mansfield Nottinghamshire NG21 0HJ Local contact Mr Will Deverill Tel: +44 (0) 333 090 8398 E-Mail: will.deverill@kordus.co.uk	Health and Hygiene - Asbestos	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene - Asbestos	B



23015

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Kordus Consulting Limited

Issue No: 001 Issue date: 02 November 2021

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248:February 2005 (Documented In-House procedure)	A, B
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 Documented In-House Procedure - using stereo-microscopy, polarised light microscopy, dispersion staining and phase contrast	A
	Sampling of bulk materials for asbestos identification	HSG 248:February 2005 Documented In-House Method	B
END			