


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>22215</b> Accredited to ISO/IEC 17025:2017	<b>Test Labs Limited</b>	
	<b>Issue No:</b> 010 <b>Issue date:</b> 20 December 2024	
	<b>Unit 22 Coningsby Road</b> <b>Bretton</b> <b>Peterborough</b> <b>PE3 8SB</b>	<b>Contact: Toni Carlton</b> <b>Tel: +44 (0)8452706690</b> <b>E-Mail: <a href="mailto:toni.carlton@testlabsuk.com">toni.carlton@testlabsuk.com</a></b> <b>Website: <a href="http://www.testlabsuk.com">www.testlabsuk.com</a></b>
<b>Testing performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
CHEMICAL DISINFECTANTS Including surface disinfectants, handwash, hand rubs	<u>Microbiological Quantitative Disinfectant Suspension Testing:</u>  Bacterial quantitative suspension test for food, industrial, domestic and institutional applications  Fungicidal or yeasticidal quantitative suspension test for food, industrial, domestic and institutional applications  Bactericidal quantitative suspension test for medical applications  Fungicidal or yeasticidal quantitative suspension test for medical applications	Documented in-house methods based on BS EN standard methods, including customer specified conditions  MIC-TP-001 based on BS EN 1276:2019  MIC-TP-003 based on BS EN 1650:2019  MIC-TP-002 based on BS EN 13727:2012 +A2:2015  MIC-TP-005 based on BS EN 13624:2021
Impregnated wipes	<u>Microbiological Quantitative Non-Porous Surface Disinfectant Testing:</u>  Bactericidal and fungicidal or yeasticidal quantitative non-porous surface test for medical applications (mechanical action employing wipes)	MIC-TP-007 based on BS EN 16615:2015



22215

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Test Labs Limited**  
**Issue No:** 010 **Issue date:** 20 December 2024

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
AUTOMATED ROOM DISINFECTION MACHINES	<u>Microbiological Quantitative Non-Porous Surface Disinfectant Testing: (Cont'd)</u>  Bactericidal, sporicidal, fungicidal and yeasticidal evaluation of automated airborne disinfection process  Bactericidal, sporicidal, fungicidal and yeasticidal evaluation of automated UV direct illumination disinfection process	Documented in-house methods based on BS EN ISO standard methods (as applicable), including customer specified conditions  MIC-TP-004 based on BS EN 17272:2020 employing test chamber sizes of 67.5m <sup>3</sup> or 11.3m <sup>3</sup>  MIC-TP-011 based on BS 8628:2022 (excluding mycobactericidal, virucidal and phagocidal activities) employing test chamber sizes of 67.5m <sup>3</sup> or 11.3m <sup>3</sup>
MEDICAL DEVICES (class 1r re-processible surgical instruments)	<u>Physical Testing:</u>  Cleaning efficacy by determination of residual protein and ATP	Documented in-house methods  TL 15883-5:2023 based on BS EN ISO 15883-5:2021 providing a simulated washer disinfectant cleaning cycle using a glassware washer (Getinge Glasswasher Ultima 810 LX) programmable to a customer-specified wash cycle, and Edinburgh soil preparation. Additional ATP assessment using 3M <sup>TM</sup> Clean-Trace <sup>TM</sup> LM1 Luminometer
MEDICAL DEVICES (class 1r re-processible surgical instruments)	Thermometric testing	TL 15883-2:2023 based on BS EN ISO 15883-2:2009
MEDICAL DEVICES (class 1r re-processible surgical instruments)	<u>Microbiological Testing:</u>  Recovery of <i>Geobacillus stearothermophilus</i> spores following sterilisation processing	TL 17665-1:2023 in accordance with BS EN ISO 17665-1:2006 Annex D, thermally processed using a Systec VX-120 autoclave at 121°C, 132°C or 134°C, with recovery onto chocolate agar or tryptone soya agar
END		