


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS MEDICAL 22998</p> <p>Accredited to ISO 15189:2012</p>	<p>PeploBio Limited</p> <p>Issue No: 001 Issue date: 04 January 2022</p>	
	<p>PeploBio Limited The Surrey Technology Centre Unit 26 40 Occam Road Guildford GU2 7YG</p>	<p>Contact: Nico Lambri E-Mail: nico@peplobio.co.uk Website: www.peplobio.co.uk</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>HUMAN BODY FLUIDS</p> <p>Combined Nose/throat swab in Viral Transport Medium</p> <p>Samples self-collected or collected by a Provider or contracted customer accredited to ISO 15189:2012 or ISO/IEC 17025:2017 and who is listed on the DHSC Approved Lists for General Population and/or Test to Release (TTR) testing</p>	<p><u>Molecular Virology examination activities for the purpose of clinical diagnosis</u></p> <p>Detection of SARS-CoV-2 virus RNA (COVID-19); specific gene sequences ORF1ab & S and N</p> <p>Detection of SARS-CoV-2 virus RNA (COVID-19); specific gene sequences N1 and E gene</p>	<p>In-house documented procedures incorporating manufacturers' instructions & DHSC requirements for private providers of Covid-19 testing, to include:</p> <ul style="list-style-type: none"> • General Population Testing (including fit-to-fly) • Test to Release (TTR) • Day 2& Day 8 for international travel <p>Automated nucleic acid extraction and Real-Time PCR using: Biorad CFX Opus platforms and ThermoFisher Quantstudio 5 using:</p> <p>ThermoFisher TaqPath Real-Time PCR Coronavirus COVID-19 (CE IVD) standard kit</p> <p>SOP0019</p> <p>Nonacus Virpath SARS Cov-2 multiplex qReal-Time-PCR kit SOP0018</p>
<p>END</p>		