


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>23137</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Cotton Mouton Diagnostics Ltd</p> <p>Issue No: 001 Issue date: 06 April 2022</p>	
	<p>8 Cae Gwyrdd Green Meadow Springs Cardiff CF15 7AB</p>	<p>Contact: Jenna Bowen Tel: +44 (0)2922 406366 E-Mail: jbowen@cm-dx.com</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
PHARMACEUTICAL LIQUID PREPARATIONS, including water for injection (WFI)	Biological Testing: Bacterial Endotoxin Testing (LAL)	<p>Documented in-house methods described below in combination with product validation methods:</p> <ol style="list-style-type: none"> 1) SOP 40 Low endotoxin recovery studies 2) SOP 51 Optimisation of LAL when endotoxin recovery is <50% or >200% <ol style="list-style-type: none"> 1) SOP 31 LAL Gel-Clot assay: WAKO Pyrostar using Pyrostar ES-F reagent 2) SOP 32 LAL turbidimetric assay using Pyrostar ES-F reagent 3) SOP 33 LAL chromogenic assay using Limulus Color KY reagent
<p>END</p>		