

Schedule of Accreditation
 issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 10641 <small>Accredited to ISO 15189: 2022</small>	<p style="text-align: center;">Inuvi Diagnostics Limited</p> <p style="text-align: center;">Issue No: 018 Issue date: 19 December 2025</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 50%;"> Unit E1 Churcham Business Park Churcham Gloucestershire GL2 8AX </td><td style="width: 50%;"> Contact: Laboratory Director Tel: +44 (0)145 222 6125 E-Mail: laboratory@inuvi.co.uk Website: www.inuvi.co.uk </td></tr> </table> <p style="text-align: center;">Testing performed at the above address only</p>		Unit E1 Churcham Business Park Churcham Gloucestershire GL2 8AX	Contact: Laboratory Director Tel: +44 (0)145 222 6125 E-Mail: laboratory@inuvi.co.uk Website: www.inuvi.co.uk
Unit E1 Churcham Business Park Churcham Gloucestershire GL2 8AX	Contact: Laboratory Director Tel: +44 (0)145 222 6125 E-Mail: laboratory@inuvi.co.uk Website: www.inuvi.co.uk			

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
HUMAN BODY FLUIDS Venous and Capillary Serum Venous and Capillary Serum, and Plasma Venous and Capillary Serum Venous and Capillary Serum Venous and Capillary Serum Venous and Capillary Serum Venous and Capillary Serum, and Plasma Venous and Capillary Serum Venous Serum Venous and Capillary Serum Venous and Capillary Serum, and Plasma Venous and Capillary Serum Venous and Capillary Serum Venous and Capillary Serum Venous Serum Venous Serum Venous Serum Venous Serum Venous Plasma Venous Serum Venous and Capillary Serum	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> <u>Quantitative analysis of:</u> Albumin ALP ALT APOA APOB Bilirubin – Total Cholesterol – Total Creatinine – Enzymatic CRP – hs Ferritin GGT HDL Iron LDH Total Protein Triglyceride Uric Acid UIBC Urea ISE (Cl, Na, K) AST Calcium Creatine Kinase Glucose Magnesium Lipoprotein A	Documented in house methods and manufacturer's instructions for: Cobas c503 analyser SOP 7.5.4 BCG IFCC IFCC Immunoturbidmetric Immunoturbidmetric Diazo Cholesterol oxidase Enzymatic Immunoturbidmetric Immunoturbidmetric IFCC Enzymatic Colourmetric IFCC Colourmetric Enzymatic Enzymatic Ferrozine Kinetic Indirect ISE IFCC NM-BAPTA UV UV Hexokinase Colourmetric Immunoturbidmetric



10641
Accredited to
ISO 15189: 2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Inuvi Diagnostics Limited

Issue No: 018 Issue date: 19 December 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd) Quantitative analysis of: Vitamin D Active Vitamin B12 Sex Hormone Binding Globulin (SHBG) Oestradiol Prolactin Follicle Stimulating Hormone (FSH) Luteinising Hormone (LH) Vitamin B12 Progesterone Free T3 Testosterone (including dilution) Folate TSH Free T4 Thyroglobulin Antibodies Thyroid Peroxidase Antibodies Total T4 ProBNP Cortisol DHEA-S PSA AMH	Documented in house methods and manufacturer's instructions for: Cobas e801 analyser SOP 7.5.4 Electrochemiluminescence Immunoassay (ECLIA)



10641
Accredited to
ISO 15189: 2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Inuvi Diagnostics Limited

Issue No: 018 Issue date: 19 December 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
Venous and Capillary Serum and Plasma	Flexible Scope Biochemical compounds	Flexible scope protocols: Biochemistry analytes and cancer markers using Roche Cobas c503 and e801. Materials or products tested: Serum, Plasma (Lithium Heparin/EDTA) (venous and/or capillary) using manufacturer validated and CE marked kits. As limited by Document ID 2.02. and the Flexible Scope List
Whole Blood (EDTA/Lithium Heparin), Venous and Capillary	HbA1c	BioRad D-100 SOP 7.7.1 HPLC



10641
Accredited to
ISO 15189: 2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Inuvi Diagnostics Limited

Issue No: 018 Issue date: 19 December 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis:</u> (cont'd) Venous & Capillary Serum Globulins LDL Cholesterol Non-HDL Cholesterol Total Iron Binding Capacity (TIBC) Transferrin Saturation Estimated Glomerular Filtration Rate (eGFR) (CK-EPI) Adjusted Calcium Free Testosterone Free Androgen Index <u>Serological examination activities for the purposes of clinical diagnosis:</u> Venous Serum Rubella IgG Antibodies Hepatitis B Surface Antibodies	By calculation
Venous Serum	Measles IgG Antibodies Mumps IgG Antibodies Varicella Zoster IgG Antibodies	Roche Cobas e801 analyser SOP 7.5.4 Dynex DS2 ELISA platform SOP 7.6.1 Novatec ELISA



10641
Accredited to
ISO 15189: 2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Inuvi Diagnostics Limited

Issue No: 018 Issue date: 19 December 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
Venous and Capillary Serum	Flexible Scope Virology/serology analytes Microbiological examinations for the purposes of clinical diagnosis	Flexible scope protocols: Virology/serology analytes using Roche Cobas e801 and Dynex DS2. Materials or products tested: Serum (venous and/or capillary) using manufacturer validated and CE marked kits. As limited by Document ID 2.02. and the Flexible Scope List
Venous Serum	Hepatitis C RNA	Transcription Mediated Amplification (TMA) using Hologic Panther, Aptima assay and inhouse method: SOP: 7.6.2
Urine, blood, serum (venous and/or capillary), swabs	Flexible Scope Microbiological RNA	Flexible scope protocols Microbiological RNA using Transcription Mediated Amplification (TMA) using Hologic Panther, Aptima assay and inhouse method. Materials or products tested: Urine, blood, serum (venous and/or capillary), swabs using manufacturer validated and CE marked kits. As limited by Document ID 2.02. and the Flexible Scope List



10641
Accredited to
ISO 15189: 2022

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Inuvi Diagnostics Limited

Issue No: 018 Issue date: 19 December 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
Whole Blood, Venous and Capillary	<p><u>Haematology examination activities for the purposes of clinical diagnosis:</u></p> <p>Full Blood Count Haemoglobin White Blood Cell Count Red Blood Count Platelet Count Haematocrit Mean Cell Volume Mean Cell Haemoglobin Mean Corpuscular haemoglobin concentration Red cell distribution width-CV Mean Platelet Volume Platelet Distribution Width Neutrophil count Lymphocyte count Monocyte count Eosinophil count Basophil count</p>	Horiba Yumizen 1500/550 Analyser using impedance, cytochemistry and absorbance SOPs: 7.2.3.3.1
END		