

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS TESTING</p> <p>20257</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Selectamark Security Systems plc</p> <p>Issue No: 002 Issue date: 20 September 2023</p>	
	<p>1 Locks Court 429 Crofton Road Locksbottom BR6 8NL</p>	<p>Contact: Jason Brown Tel: +44 (0)1689 860757 E-Mail: jason.brown@selectamark.co.uk Website: www.selectadna.co.uk</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>SELECTAMARK PRODUCTS</p> <p>Any solid material</p>	<p>Forensic Testing</p> <p><u>Forensic Analysis</u></p> <p>Searching for the presence of SelectaDNA products</p> <p>Recovery of SelectaDNA products and preparation for subsequent DNA analysis</p> <p>Analysis of SelectaDNA products</p>	<p>The organisation has demonstrated compliance to the Forensic Science Regulator Code of Practice in relation to the Forensic Activities listed below.</p> <p>Documented in-house method (SOP.03) using visual examination and UV light source 365nm.</p> <p>Documented in-house method (SOP.03) using:</p> <ul style="list-style-type: none"> - Cutting - Swabs and Swabbing - Scraping <p>Documented in-house methods:</p> <ul style="list-style-type: none"> - Extraction (SOP.04, SOP.05) - Amplification (SOP.07) - Electrophoresis (SOP.08) - Sequencing (SOP.09) - Evaluation of Analysis Results (SOP.10)
<p>END</p>		