


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 23524 Accredited to ISO/IEC 17025:2017	Vacuum and Atmosphere Services Ltd	
	Issue No: 002 Issue date: 16 April 2025	
	Unit 30, Park Rose Industrial Estate Middlemore Road Birmingham B66 2DZ	Contact: Martin Perry Tel: +44 (0)121 544 4385 E-Mail: M.Perry@vacat.co.uk Website: www.vacat.co.uk
Calibration performed by the Organisation at the locations specified below		

Site activities performed away from the location listed above:

Location details	Activity
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Temperature Electrical



23524
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Vacuum and Atmosphere Services Ltd
Issue No: 002 Issue date: 16 April 2025

Site activities performed away from the location listed above:

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	location
TEMPERATURE Temperature controlled, ovens, environmental chambers and furnaces (inclusive of associated indicators, controllers and recorders, all with sensors)	100 °C to 600 °C 600 °C to 1100 °C 1100 °C to 1300 °C	1.4 °C 2.2 °C 3.4 °C	Multipoint calibration also referred to as spatial temperature surveying or mapping	Site
ELECTRICAL Thermocouple simulators and indicators, calibration by electrical simulation Noble metal thermocouples Type R, Type S Base metal thermocouples Type N, Type K, Type J	400 °C to 1760 °C 0 °C to 1300 °C	0.51 °C 0.39 °C	excludes cold junction compensation	
Measurement DC Current DC Volts	0 mA to 20 mA 0 V to 10 V	0.0033 mA 0.0016 V		
END				



23524
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Vacuum and Atmosphere Services Ltd
Issue No: 002 Issue date: 16 April 2025

Site activities performed away from the location listed above:

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$